



AC 114

## CERTYFIKAT BADANIA TYPU UE (MODUŁ B) EU TYPE-EXAMINATION CERTIFICATE (MODULE B)

Nr  
No. CW/PPER/28/12/2020

### ZAŚWIADCZA SIĘ,

że Polski Rejestr Statków S.A. (PRS) przeprowadził procedurę badania typu wymienionego niżej wyrobu i stwierdził jego zgodność z wymaganiami określonymi w załączniku V do Rozporządzenia Parlamentu Europejskiego i Rady (UE) 2016/425 (PPE) w sprawie środków ochrony indywidualnej oraz uchylecia dyrektywy Rady 89/686/EWG, ze zmianami.

### THIS IS TO CERTIFY

that Polski Rejestr Statków S.A. (PRS) did undertake the EU type-examination procedure for the product identified below which was found to be in compliance with the requirements of Annex V to the Regulation (EU) 2016/425 of the European Parliament and of the Council of 9 March 2016 on personal protective equipment and repealing Council Directive 89/686/EEC, as amended.

Wnioskodawca  
Applicant Baoding Yinhong Yuhe medical device manufacturing Co., Ltd.  
Nanlongshan village, Dawangdian Industrial Park,  
Xushui District, Baoding City, Hebei Province, China.

Producent  
Manufacturer Baoding Yinhong Yuhe medical device manufacturing Co., Ltd.  
Nanlongshan village, Dawangdian Industrial Park,  
Xushui District, Baoding City, Hebei Province, China.

Typ wyrobu  
Product type **Sprzęt ochrony układu oddechowego. Półmaski filtrujące do ochrony przed cząstkami.**  
*Respiratory protective devices. Filtering half masks to protect against particles.*

Opis wyrobu  
Product description **Półmaska filtrująca, model: YH/9903 (klasa FFP3 NR).**  
*Filtering half mask, Model: YH/9903 (class FFP3 NR).*

Zastosowane normy  
Specified standards PN-EN 149+A1:2010  
EN 149:2001+A1:2009

Niniejszy certyfikat pozostaje ważny do czasu unieważnienia przy zachowaniu warunków uznania (patrz str. 2).  
*This certificate remains valid unless cancelled or revoked, provided the approval conditions (see page 2) are complied with.*

Data ważności  
Expiry date 2025-12-08



Zastępca Dyrektora Biuru Certyfikacji  
Certification Division Deputy Director

Przemysław Gałka

Gdańsk, 2020-12-09



Nr jednostki notyfikowanej  
No. of notified body  
**1463**

Polski Rejestr Statków S.A.  
al. Gen. Józefa Hallera 126  
80-416 Gdańsk, Poland

tel. (+48) (58) 346 17 00  
fax (+48) (58) 341 77 69  
e-mail: dc@prs.pl  
www: <http://www.prs.pl/>

Wykaz dokumentacji  
List of documents

1. Instrukcja użytkowania - zatwierdzona przez PRS S.A. dnia 2020-12-03.
2. Ocena ryzyka - zatwierdzona przez PRS S.A. dnia 2020-12-03.
3. Dokumentacja techniczna „Półmaski filtrującej, model: YH/9903” - zatwierdzony przez PRS S.A. dnia 2020-12-03.
4. Raport z badań nr JKf20032192R1 wydany przez Zhejiang Academy of Science and Technology for Inspection & Quarantine (Technology Center of Hangzhou Customs District/Zhejiang Lead Product Technical Co., Ltd.) z akredytacją CNAS L0354 z dnia 2020-12-08.
5. Sprawozdanie z przeglądu PRS S.A. nr CW/MoK/PPER/277/2020 z dnia 2020-12-09.

1. *Instruction of use - approved by PRS S.A. on 2020-12-03.*
2. *Risk analysis - approved by PRS S.A. on 2020-12-03.*
3. *Technical documentation "Filtering half mask, Model: YH/9903" - approved by PRS S.A. on 2020-12-03.*
4. *Test report No. JKf20032192R1 issued by Zhejiang Academy of Science and Technology for Inspection & Quarantine (Technology Center of Hangzhou Customs District/Zhejiang Lead Product Technical Co., Ltd.) with CNAS accreditation no. L0354 dated on 2020-12-08.*
5. *PRS S.A. Survey Report No. CW/MoK/PPER/277/2020 dated on 2020-12-09.*

Miejsca produkcji  
(inne niż podane na stronie 1)  
Places of production  
(different than given on page 1)

- - -

Ograniczenia uznania  
Approval limitations

1. Dane techniczne:
  - a) półmaska filtrująca z regulowanym klipsem na nos,
  - b) klips na nos montowany wewnątrz półmaski filtrującej,
  - c) półmaska filtrująca wykonana z 5 warstwowej włókniny z filtrem z tkaniny,
  - d) półmaska filtrująca wyposażona w zauszniki,
  - e) półmaska filtrująca bez zaworu,
  - f) wymiary: 100 mm ± 2 mm x 153 mm ± 3 mm,
  - g) docelowa grupa użytkowa: dorośli dla obu płci,
  - h) kolory:

półmaska filtrująca	zagłowie	klips na nos	zawór
biała	białe	n / d	n / d

2. Półmaska filtrująca przeznaczona do jednorazowego użytku.
3. Dokumentacja techniczna zatwierdzona w języku angielskim.
4. Produkt ten nie może być stosowany jako maska przeciwgazowa w środowisku toksycznym.
5. Półmaska filtrująca nie powinna być używana w środowisku o stężeniu tlenu poniżej 19.5 %.
6. Półmaska filtrująca nie jest przeznaczona do użytkowania medycznego i chirurgicznego.

## 1. Specifications:

- a) filtering half mask with adjustable nose clip,
- b) nose clip mounted inside the filtering half mask,
- c) filtering half mask made with 5 layers non-woven fabric with melt-blown fabric filter,
- d) filtering half mask with ear loops,
- e) filtering half mask without valve,
- f) size: 100 mm ± 2 mm x 153 mm ± 3 mm,
- g) target group: unisex,
- h) colors:

filtering half mask	head harness	nose clip	valve
white	white	NA	NA

2. *Filtering half mask shall not be used for more than one shift.*
3. *Technical documentation approved in English.*
4. *This product can not be used as a gas mask in a toxic environment.*
5. *Filtering half mask should not be used in an environment with oxygen contents less than 19.5%.*
6. *Filtering half mask can not be used for medical and surgical purposes.*

Warunki uznania  
Approval conditions

- 1 Niniejszy certyfikat straci ważność po wprowadzeniu zmian lub modyfikacji w wyrobie bez uprzedniego uzgodnienia z PRS.  
*This certificate becomes invalid after changes or modifications to the product without prior agreement with PRS.*
- 2 Znak zgodności może być umieszczony na uznanym wyrobie oraz może być wystawiona deklaracja zgodności tylko pod warunkiem, że łącznie z badaniem typu UE zostanie przeprowadzona ocena zgodności produkcji pod nadzorem jednostki notyfikowanej, według załącznika VII lub VIII wymienionego wyżej rozporządzenia.  
*The Mark of Conformity may only be affixed to the above type approved product and a manufacturer's Declaration of Conformity issued provided the production is assessed under surveillance of a notified body according to Annex VII or VIII of the a/m Regulation.*



AC 114

**CERTYFIKAT ZGODNOŚCI Z TYPEM W OPARCIU O WEWNĘTRZNA KONTROLĘ  
PRODUKCJI ORAZ NADZOROWANE KONTROLE PRODUKTU  
W LOSOWYCH ODSTĘPACH CZASU (Moduł C2)**

**CONFORMITY TO TYPE CERTIFICATE BASED ON INTERNAL PRODUCTION CONTROL  
PLUS SUPERVISED PRODUCT CHECKS AT RANDOM INTERVALS (Module C2)**

Nr CW/PPER/66/12/2020 Okres objęty certyfikatem 2020-12-19 – 2021-12-18  
No. Period covered by the certificate

Dokumenty odniesienia: Rozporządzenie UE 2016/425 dotyczące środków ochrony indywidualnej (PPE), załącznik VII  
General reference documents: Regulation (EU) 2016/425 on personal protective equipment (PPE), Annex VII

Posiadacz certyfikatu Baoding Yinhong Yuhe medical device manufacturing Co., Ltd.  
Certificate holder Nanlongshan village, Dawangdian Industrial Park,  
Xushui District, Baoding City, Hebei Province, China.

Wyrób Product	Certyfikat badania typu UE EU Type-examination certificate	Normy zharmonizowane/Specyfikacje Harmonised standards/Specifications
<b>Półmaska filtrująca, model: YH/9903 (klasa FFP3 NR).</b> <i>Filtering half mask, Model: YH/9903 (class FFP3 NR).</i>	CW/PPER/28/12/2020	PN-EN 149+A1:2010 EN 149:2001+A1:2009

**A Roczna ocena zgodności wyrobów z normą/specyfikacją i badanym typem**

**Annual assessment of products compliance with standard/specification and type-examined**

- 1 Miejsca i daty wizyt  
Visit locations and dates Baoding Yinhong Yuhe medical device manufacturing Co., Ltd.
- 2a Wyboru dokonał (imię, nazwisko)  
Selection carried out by (Name) Mirosław Klimek  
Związek z jednostką notyfikowaną  
Relationship to notified body Ekspert Biura Certyfikacji Wyrobów i Osób  
Products and Persons Certification Bureau Expert
- 2b Przedstawiciel firmy (imię, nazwisko)  
Company representative (Name) Liu Henghao  
Stanowisko  
Position -
- 3 Związek pomiędzy wizytowaną firmą a posiadaczem certyfikatu badania typu UE  
Relationship of company visited to EU type-examination certificate holder
- Posiadacz certyfikatu  
Certificate holder  Miejsce produkcji  
Production site  Inne miejsce produkcji  
Secondary production site  Importer  Dystrybutor  
Distributor
- Sprzedaż detaliczna  
Retail outlet  Europejskie biuro firmy  
European office of the company  Inny:  
Other:
- Wykaz środków ochrony indywidualnej  
List of personal protection equipment  Dostępny  
Available  Niedostępny  
Not available
- Wybór próbek  
Sample selection  Wybrano – Nr egz./partii:  
Selected – lot/batch No. YH/9903/2020/11/3  Nie wybrano  
Not selected
- 4 Wybór próbek  
Sample selection  Prawidłowy  
Correct  Nieprawidłowy  
Incorrect Wyniki badań  
Result of tests  Pozytywne  
Positive  Negatywne  
Negative
- 5 Wybór próbek i badania wykazały zgodność z przywołanymi normami/specyfikacjami i badanym typem  
Sample selection and testing demonstrated compliance with the reference standards/specifications and type-examined  Tak  
Yes  Nie  
No



Nr jednostki notyfikowanej  
No. of notified body

1463

Polski Rejestr Statków S.A.  
al. Gen. Józefa Hallera 126  
80-416 Gdańsk, Poland

tel. (+48) (58) 346 17 00  
fax (+48) (58) 346 03 92  
e-mail: mailbox@prs.pl  
www: http://www.prs.pl/

**B Roczna ocena niejednorodności produkcji**  
**Annual assessment of production non-homogeneity**

1 Zastosowana metoda przy dokonaniu oceny  
 Method employed to perform assessment

- Inspekcja procesu produkcyjnego i zapisów z prób  
 On-site review of production and test records
- Audit kontroli procesu produkcyjnego  
 On-site audit of production control
- Ocena niejednorodności produkcji poprzez ocenę jednej dużej próbki  
 Production non-homogeneity assessed by selection of a single, large sample
- Ocena niejednorodności produkcji poprzez ocenę próbek w ciągu roku  
 Production non-homogeneity assessed by assessment of samples throughout the year

2a Ocenę przeprowadził (imię, nazwisko) ----  
 Assessment carried out by (Name)  
 Związek z jednostką notyfikowaną ----  
 Relationship to notified body

2b Przedstawiciel firmy (imię, nazwisko) ----  
 Company representative (Name)  
 Stanowisko ----  
 Position

3 Na podstawie przeprowadzonej oceny stwierdzono, że proces produkcyjny jest jednorodny  Tak  Nie  
 On the basis of the assessment, it has been concluded the production is homogeneous Yes No

**C Podsumowanie**  
**Conclusion**

Uzasadnienie niezgodności  
 Justification of non-conformities

Nie było żadnych niezgodności / There were no non-conformities.

Wnioski jednostki notyfikowanej  
 Conclusions of notified body

Środek ochrony osobistej jest kompatybilny z typem określonym w certyfikacie badania typu UE.  
 Personal protective equipment is compatible with the type defined in the EC type-examination certificate.

Uwagi  
 Remarks

1. Półmaska filtrująca przeznaczona do jednorazowego użytku.
  2. Dokumentacja techniczna zatwierdzona w języku angielskim.
  3. Produkt ten nie może być stosowany jako maska przeciwgazowa w środowisku toksycznym.
  4. Półmaska filtrująca nie jest przeznaczona do użytkowania medycznego i chirurgicznego.
  5. Półmaska filtrująca nie powinna być używana w środowisku o stężeniu tlenu poniżej 19.5 %.
1. Filtering half mask shall not be used for more than one shift.
  2. Technical documentation approved in English.
  3. This product can not be used as a gas mask in a toxic environment.
  4. Filtering half mask can not be used for medical and surgical purposes.
  5. Filtering half mask should not be used in an environment with oxygen contents less than 19.5%.

**D Załączniki**  
**Attachments**

Sprawozdania z wizyty Nr CW/MoK/PPER/304/2020 z dnia/dated on 2020-12-19.  
 Visit reports No.

Sprawozdania z badań Nr Raport z badań nr CL/WBO/152/2020 wydany przez Laboratorium Badawcze PRS S.A. w dniu  
 Test reports No. 2020-12-19.  
 Test report no. CL/WBO/152/2020 issue by PRS. S.A. Testing Laboratory dated on 2020-12-19.

Ogólna ocena z rocznego nadzoru  
 Overall assessment of the annual surveillance

Pozytywna  
 Positive

Negatywna  
 Negative




Zastępca Dyrektora Pionu Certyfikacji  
 Certification Division Deputy Director

*[Signature]*  
 Przemysław Gałka

Gdańsk, 2020-12-19

# TEST REPORT



**Report No.: JKF20032192R1**

**Applicant : Baoding Yinhong Yuhe medical device manufacturing  
Co., Ltd**

**Zhejiang Academy of Science and Technology for Inspection and Quarantine**

Add: No. 398, Jianshe 3 Road, Xiaoshan District, Hangzhou, Zhejiang, China

Tel: +86 0571 8352 7187/185/193 Website: [www.zaiq.org.cn](http://www.zaiq.org.cn)



Report No.: JKF20032192R1

Report date: 2020-12-08

The information are provided by client(applicant):				
Sample Information	Sample Name:	filtering half mask		
	Style No.:	YH/9903		
	Brand:	YINHONYUHE		
Customer Information	Applicant:	Baoding Yinhong Yuhe medical device manufacturing Co., Ltd		
	Address:	Nanlongshan village, Dawangdian Industrial Park, Xushui District, Baoding City, Hebei Province, China		
	Manufacturer:	Baoding Yinhong Yuhe medical device manufacturing Co., Ltd		
	Manufacturer address:	Nanlongshan village, Dawangdian Industrial Park, Xushui District, Baoding City, Hebei Province, China		
The information are confirmed by testing organization:				
Test Information	Date of sample received:	2020-11-24	Testing period:	2020-11-24 to 2020-12-07
	Quantity:	70 Pieces		
	Sample description:	White mask		
	Basis of judgment:	EN 149:2001+A1:2009 FFP3 NR Respiratory protective devices—Filtering half masks to protect against particles —Requirements, testing, marking		
Test Conclusion	The items tested meet the requirements of EN 149:2001+A1:2009 FFP3 NR			
Test Result	Please refer to next pages.			
Remark	This report (which has modified Sample photo) is to replace the original report (report number JKF20032192 issued on 2020-12-07), the original report also void.			

Edit:

叶义文

Ye yiwen

Sign:

Zhao dong

\*\*\* End of this page\*\*\*



## Test Results:

### Clause 7.5 Material

(EN 149:2001+A1:2009 Clause 8.2 & 8.3.1 & 8.3.2)

Requirement	Results	Rating
<p>Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used.</p> <p>After undergoing the conditioning described in 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps.</p> <p>When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse.</p> <p>Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer.</p>	Comply	Pass

### Clause 7.6 Cleaning and disinfecting

(EN 149:2001+A1:2009 Clause 8.4 & 8.5 & 8.11)

Requirement	Results	Rating
<p>If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer.</p> <p>With reference to 7.9.2, after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class.</p>	Not applicable (Not designed to be re-usable)	N/A

### Clause 7.7 Practical performance

(EN 149:2001+A1:2009 Clause 8.4)

Requirement	Results	Rating
<p>The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard.</p>	No imperfections	Pass

### Clause 7.8 Finish of parts

(EN 149:2001+A1:2009 Clause 8.2)

Requirement	Results	Rating
<p>Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs.</p>	No sharp edges or burrs	Pass

**Clause 7.9.1 Total inward leakage**

(EN 149:2001+A1:2009 Clause 8.5)

Requirement	Results	Rating
For particle filtering half masks fitted in accordance with the manufacturer's information, at least 46 out of the 50 individual exercise results (i.e. 10 subjects x 5 exercises) for total inward leakage shall be not greater than: 25% for FFP1, 11% for FFP2, 5% for FFP3 and, in addition, at least 8 out of the 10 individual wearer arithmetic means for the total inward leakage shall be not greater than: 22% for FFP1, 8% for FFP2, 2% for FFP3	50 out of the 50 individual exercise $\leq$ 5% 8 out of the 10 individual wearer arithmetic means $\leq$ 2%	Pass

Table 7.9.1-A Inward leakage test data

Subject	Sample No.	Condition	Walk (%)	Head side/side (%)	Head up/down (%)	Talk (%)	Walk (%)	Mean (%)
CQQ	1	As received	1.778	1.731	1.795	2.909	1.855	2.013
WLJ	2		1.785	1.778	1.806	2.867	1.874	2.022
WG	3		1.695	1.782	1.792	2.718	1.858	1.969
ZJH	4		1.673	1.727	1.740	2.627	1.816	1.916
TLB	5		1.762	1.679	1.696	2.528	1.780	1.889
ZMY	6	Temperature conditioned	1.801	1.671	1.712	2.454	1.773	1.882
LJF	7		1.722	1.678	1.742	2.847	1.793	1.956
HML	8		1.627	1.677	1.700	2.651	1.699	1.871
RK	9		1.605	1.810	1.736	2.457	1.802	1.882
ZD	10		1.677	1.692	1.765	2.688	1.825	1.929

Table 7.9.1-B Facial dimensions

Subject	Face Length (mm)	Face Width (mm)	Face Depth (mm)	Mouth Width (mm)
CQQ	136	167	125	65
WLJ	132	159	110	60
WG	120	152	109	57
ZJH	122	150	104	50
TLB	125	152	111	57
ZMY	137	150	120	60
LJF	125	135	90	55
HML	124	130	115	55
RK	112	161	146	50
ZD	116	160	115	55



**Clause 7.9.2 Penetration of filter material**

(EN 149:2001+A1:2009 Clause 8.11 & EN 13274-7:2019)

Requirement			Results	Rating
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table.			Detail refer to Table 7.9.2	Pass
Classification	Sodium chloride test 95 L/min	Paraffin oil test 95 L/min		
FFP1	≤20%	≤20%		
FFP2	≤6%	≤6%		
FFP3	≤1%	≤1%		

Table 7.9.2 Penetration of filter material

Aerosol	Condition	Sample No.	Penetration (%)
Sodium chloride test	As received	11	0.005
		12	0.007
		13	0.006
	Simulated wearing treatment	14	0.007
		15	0.004
		16	0.005
	Mechanical strength+ Temperature conditioned	17	0.012
		18	0.017
		19	0.020
Paraffin oil test	As received	20	0.033
		21	0.025
		22	0.024
	Simulated wearing treatment	23	0.013
		24	0.019
		25	0.022
	Mechanical strength+ Temperature conditioned	26	0.630
		27	0.527
		28	0.593
Flow conditioning: single filter: 95.0 L/min			

**Clause 7.10 Compatibility with skin**

(EN 149:2001+A1:2009 Clause 8.4 & 8.5)

Requirement	Results	Rating
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health.	No irritation or any other adverse effect to health	Pass

**Clause 7.11 Flammability**

(EN 149:2001+A1:2009 Clause 8.6)

Requirement	Results	Rating
When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5s after removal from the flame.	Detail refer to Table 7.11	Pass

Table 7.11 Flammability

Condition	Sample No.	Result
As received	29	Not burn
	30	Not burn
Temperature conditioned	31	Not burn
	32	Not burn

**Clause 7.12 Carbon dioxide content of the inhalation air**

(EN 149:2001+A1:2009 Clause 8.7)

Requirement	Results	Rating
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0 % (by volume).	Detail refer to Table 7.12	Pass

Table 7.12 Carbon dioxide content of the inhalation air

Condition	Sample No.	Result (%)
As received	33	0.69
	34	0.69
	35	0.70
		Mean value: 0.69

**Clause 7.13 Head harness**

(EN 149:2001+A1:2009 Clause 8.4 & 8.5)

Requirement	Results	Rating
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily. The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device.	Comply	Pass

**Clause 7.14 Field of vision**

(EN 149:2001+A1:2009 Clause 8.4)

Requirement	Results	Rating
The field of vision is acceptable if determined so in practical performance tests.	Comply	Pass

**Clause 7.15 Exhalation valve**

(EN 149:2001+A1:2009 Clause 8.2 & 8.9.1 & 8.3.4 & 8.8)

Requirement	Results	Rating
<p>A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations.</p> <p>If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9.</p> <p>Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 L/min over a period of 30 s.</p> <p>When the exhalation valve housing is attached to the faceblank, it shall withstand axially a tensile force of 10 N applied for 10 s.</p>	Not applicable (No exhalation valve)	N/A

**Clause 7.16 Breathing resistance**

(EN 149:2001+A1:2009 Clause 8.9)

Requirement	Results	Rating																						
<p>The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table.</p> <table border="1" data-bbox="263 1064 1021 1317"> <thead> <tr> <th rowspan="3">Classification</th> <th colspan="3">Maximum permitted resistance (mbar)</th> </tr> <tr> <th colspan="2">Inhalation</th> <th>Exhalation</th> </tr> <tr> <th>30L/min</th> <th>95L/min</th> <th>160L/min</th> </tr> </thead> <tbody> <tr> <td>FFP1</td> <td>0.6</td> <td>2.1</td> <td>3.0</td> </tr> <tr> <td>FFP2</td> <td>0.7</td> <td>2.4</td> <td>3.0</td> </tr> <tr> <td>FFP3</td> <td>1.0</td> <td>3.0</td> <td>3.0</td> </tr> </tbody> </table>	Classification	Maximum permitted resistance (mbar)			Inhalation		Exhalation	30L/min	95L/min	160L/min	FFP1	0.6	2.1	3.0	FFP2	0.7	2.4	3.0	FFP3	1.0	3.0	3.0	Detail refer to Table 7.16	Pass
Classification		Maximum permitted resistance (mbar)																						
		Inhalation		Exhalation																				
	30L/min	95L/min	160L/min																					
FFP1	0.6	2.1	3.0																					
FFP2	0.7	2.4	3.0																					
FFP3	1.0	3.0	3.0																					

Table 7.16 Breathing resistance (mbar)

Test item	Condition	Sample No.	A	B	C	D	E
Inhalation (30 L/min)	As received	36	0.50	0.51	0.51	0.51	0.50
		37	0.49	0.50	0.50	0.49	0.49
		38	0.49	0.50	0.49	0.50	0.50
	Simulated wearing treatment	39	0.51	0.51	0.50	0.51	0.51
		40	0.51	0.52	0.51	0.52	0.52
		41	0.50	0.51	0.51	0.51	0.50
	Temperature conditioned	42	0.45	0.46	0.46	0.45	0.45
		43	0.44	0.45	0.45	0.45	0.44
		44	0.47	0.46	0.47	0.46	0.46

Test item	Condition	Sample No.	A	B	C	D	E
Inhalation (95 L/min)	As received	36	1.87	1.88	1.87	1.86	1.88
		37	1.85	1.86	1.87	1.85	1.85
		38	1.84	1.86	1.85	1.86	1.85
	Simulated wearing treatment	39	1.89	1.88	1.87	1.89	1.89
		40	1.90	1.91	1.92	1.94	1.92
		41	1.90	1.88	1.89	1.89	1.88
	Temperature conditioned	42	1.79	1.77	1.79	1.77	1.77
		43	1.76	1.74	1.74	1.75	1.74
		44	1.81	1.80	1.82	1.83	1.81
Exhalation (160 L/min)	As received	36	2.65	2.64	2.66	2.65	2.65
		37	2.64	2.62	2.64	2.63	2.64
		38	2.63	2.64	2.65	2.63	2.63
	Simulated wearing treatment	39	2.67	2.65	2.66	2.67	2.67
		40	2.69	2.70	2.69	2.67	2.68
		41	2.68	2.66	2.67	2.65	2.66
	Temperature conditioned	42	2.48	2.45	2.47	2.47	2.45
		43	2.45	2.47	2.46	2.45	2.46
		44	2.49	2.51	2.50	2.51	2.49

A: facing directly ahead; B: facing vertically upwards; C: facing vertically downwards; D: lying on the left side; E: lying on the right side

### Clause 7.17 Clogging

(EN 149:2001+A1:2009 Clause 8.9 & 8.10)

Requirement	Results	Rating
<p><b>7.17.2 Breathing resistance:</b></p> <p><b>7.17.2.1 Valved particle filtering half masks</b> After clogging the inhalation resistances shall not exceed FFP1:4mbar, FFP2:5mbar, FFP3:7mbar at 95 L/min continuous flow; The exhalation resistance shall not exceed 3mbar at 160 L/min continuous flow.</p> <p><b>7.17.2.2 Valveless particle filtering half masks</b> After clogging the inhalation and exhalation resistances shall not exceed FFP1:3mbar, FFP2:4mbar, FFP3:5mbar at 95 L/min continuous flow.</p> <p><b>7.17.3 Penetration of filter material:</b> All types (valved and valveless) of particle filtering half masks claimed to meet the clogging requirement shall also meet the requirements given in 7.9.2, for the Penetration test according to EN 13274-7, after the clogging treatment.</p>	Optional for single shift device only	Not required

### Clause 7.18 Demountable parts

(EN 149:2001+A1:2009 Clause 8.2)

Requirement	Results	Rating
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand.	Comply	Pass

### Sample photo



**\*\*\* End of Report\*\*\***

## STATEMENT

1. Our organization guarantees impartiality, independence and honesty of inspection, and is responsible for the content of report, except for the information provided by the client. The client shall not use the test results for improper publicity without authorization.
2. Our organization shall not be responsible for the authenticity of the information provided by the client, nor shall bear the risks arising in the process of sample delivery. Test result is only responsible for the sample.
3. This report is invalid without the dedicated seal for inspection and testing report and the paging seal.
4. This report is invalid without the signature of the approver (authorized signatory).
5. Test report is invalid if altered.
6. The duplicate report without the "dedicated seal for inspection and testing" of the institution is invalid.
7. Each page of the report is an integral part of the report. Our organization shall not be responsible for any misunderstanding or consequences arising from the improper use of the test report by the user.
8. Without the CMA seal, the report is invalid for social certification.

Test institute: Zhejiang Academy of Science and Technology for Inspection and Quarantine

Addr: No. 398, Jianshe 3 Road, Xiaoshan District, Hangzhou, Zhejiang, China

Tel: +86 0571 8352 7187/185/193

Website: [www.zaiq.org.cn](http://www.zaiq.org.cn)